

A methodology to standardize the evaluation of root canal instrumentation using cone beam computed tomography

Cláudia Bohrer Flores¹, Patrícia Machado¹, Francisco Montagner²,
Brenda Paula Figueiredo de Almeida Gomes³, Gustavo Nogara Dotto⁴, Marcia da Silva Schmitz⁴

¹DDS, MSc student, Federal University of Santa Maria, Brazil

²DDS, MSc, PhD, Adjunct Professor, Endodontic Division, Department of Conservative Dentistry, Federal University of Rio Grande do Sul, Brazil

³DDS, MSc, PhD, Head Professor, Endodontic Division, Department of Conservative Dentistry, Piracicaba Dental School, University of Campinas, Brazil

⁴DDS, MSc, PhD, Adjunct Professor, Endodontic Division, Department of Stomatology, Federal University of Santa Maria, Brazil

Abstract

The evaluation of root canal instrumentation is important to observe the action of endodontic instruments in the root canal walls. **Aim:** This study introduces a method to standardize the acquisition of images before and after preparing root canals by using cone-beam computed tomography (CBCT). **Methods:** Sixteen mandibular molars were included in acrylic resin blocks. Samples were inserted in a stable wood box, which was filled with plaster and served as a guide to reinsert the samples. The apparatus was used for the CBCT examination before and after cervical flaring of root canals. The software IcatVision[®] was used to equalize the images before and after instrumentation with two computers operating at the same time. The measurements between root canal center and the furcation area were determined. The statistical analysis was performed using the t-test for paired samples ($\alpha=0.05$). **Results:** The values for dentin thickness in the risk zone before and after root canal flaring with Gates-Gliden drills were 1.096 (± 0.27) mm and 0.742 (± 0.24) ($p<0.01$), respectively. **Conclusions:** The proposed method assures the same positioning of the samples before and after root canal preparation. It is extremely important, as any mesiodistal or buccolingual movement can produce a different topogram for comparison procedures.

Keywords: cone-beam computed tomography, anatomy cross-sectional, image processing computer-assisted, endodontics.

Introduction

The evaluation of root canal instrumentation is necessary to determine the action of the instruments on the original canal shape¹⁻² and to observe if the principles of canal preparation had been followed³. The thin area in the root canal wall (risk zone), vulnerable to stripping by injudicious filing⁴, has been analyzed using microscopic analyses⁵⁻⁶, silicone impressions⁷, muffle system⁸⁻¹⁴, Endodontic Cube¹⁵⁻¹⁶, multi-slice computed tomography¹⁷ and cone-beam computed tomography (CBCT)¹⁸.

CBCT evaluation has shown that the exact location and anatomy of the root canal system can be assessed¹⁹⁻²¹ and it has been validated as a tool to explore root canal anatomy²². Furthermore, it has been successfully used for measurements

Received for publication: November 25, 2011

Accepted: April 25, 2012

Correspondence to:

Francisco Montagner

Universidade Federal do Rio Grande do Sul,
Faculdade de Odontologia
Rua Ramiro Barcelos 2492, CEP: 90035-003
Bairro Bom Fim - Porto Alegre, RS - Brazil
Phone/Fax: +55 51 33085430
E-mail: francisco.montagner@ufrgs.br

before and after instrumentation of root canals and for determining the amount of dentin removed during cleaning and shaping of root canals¹⁸.

However, a method has not been proposed to standardize the positioning of samples for obtaining images or topograms before and after root canal preparation by CBCT. It is extremely important because any displacement of the sample in both buccolingual and mesiodistal directions will not provide pre- and post-instrumentation image superimposition. Therefore, the aim of this study was to develop a methodology to standardize specimen positioning before and after image acquisition by CBCT.

Material and methods

This study was approved by the Ethics Committee of the Santa Maria University, Santa Maria, RS, Brazil.

Sixteen extracted periodontally involved mandibular molars with complete root formation were radiographed and stored in 0.1% thymol solution at 9° C. Teeth were placed under tap water for 12 h prior to use to eliminate residues of the storage solution. The teeth that exhibited previous endodontic manipulation, internal or external resorption, were excluded from the study.

Coronal access was performed using a #1014 diamond bur (KG Sorensen Cotia, SP, Brazil) followed by a Endo-Z drill (Dentsply, Maillefer, Ballaigues, Switzerland). The root canal of each tooth was flushed with 2.5% NaOCl (Manipulation Pharmacy, Nova Derme, Santa Maria, RS, Brazil) and explored using a size 10 K-ûle (Dentsply Maillefer) until the apical foramen was reached. Root canal length was determining through the visualization of the #10 file tip.

The entire root apex was covered with colorless nail polish (Colorama, Procosa Produtos de Beleza Ltda, São Paulo, SP, Brazil), to avoid the penetration of acrylic resin into the root canal system. A plastic box (1.5 cm x 1.5 cm x 1.5 cm) was used as guide where the teeth were inserted. The plastic boxes were lubricated with solid Vaseline (Rioquímica, São José do Rio Preto, SP, Brazil). Each tooth was placed inside the plastic box and the self-curing acrylic resin was inserted to position the teeth, without covering the crown (Figures 1A and 1B). After 24 h, the acrylic resin was polished (Arotec, Cotia, SP, Brazil) with 200-, 400- and 600-grit abrasive paper for 30 s each (Alcar abrasivos, Vinhedo, SP, Brazil). The specimen's surface were covered with a thin layer of separating medium (Cel-Lac; S.S. White, Rio de Janeiro, RJ, Brazil).

A 10 cm x 10 cm x 3 cm wood box was made to serve as a template to insert the type III plaster (Polidental Ltda., Cotia, SP, Brazil) (Figure 1C). The internal surfaces of the wood box were covered with plastic film (Wyda, Sorocaba, SP, Brazil) to favor cast dislodgement after setting period. The plaster was then prepared and inserted in the box. The specimens were placed inside the apparatus as shown in Figure 1D. After the plaster setting, the specimens were removed. Both specimens and the spaces that were formed in the casts received a code to identify them (Figure 1E).

After 24 h, the specimens were replaced in the cast




Fig. 1. Sample preparation and imaging acquisition: A) Plastic cube; B) Specimen embedded in acrylic resin; C) Wood mould; D) Specimen embedded in gypsum; E) Numbering of specimens in the corresponding depression; F) Mould of gypsum aligned through the rotation tool.

according to the codes and 3-dimensional CBCT images (GX CB-500 POWERED BY i-CAT) were obtained with exposure time of 26 s, operating at 120 Kvp and 5 mA. CBCT imaging was performed with the 0.25-mm voxel size. The reconstruction of the sections was performed.

Cervical instrumentation of the root canals was performed with size 1 and 2 Gates-Glidden I (Dentsply, Maillefer) under irrigation with 2.5% NaOCl.

The specimens were then replaced on the apparatus and a new CBCT imaging acquisition was performed as previously described. IcatVision® software (Dental Imaging System, Salt Lake City, UT, USA) was used to equalize the images before and after instrumentation. The images were aligned, observing the axial vision, through the rotation tool (Figure 1F). The MPR Screen was selected for measuring. The “zoom” tool was applied to allow a better visualization of the tooth. The vertical (blue and red) and horizontal (green) bars were used as reference to align the images. The tool “distance” (on coronal section) was employed to determine the measure from the highest point of the furcation area up to 2 mm apically. Then, the horizontal bar was adjusted 2 mm from furcation

area (Figure 2A), generating an image in the axial section. The distance from the center of the root canal to the furcation area were determined (Figures 2B and 2C).

Statistical analysis was carried out in the BioEstat 5.0 (Fundação Mamirauá, Belém, PA, Brazil), and the t-test for paired samples was used for the comparison between groups ($p < 0.01$).




Fig. 2. Data acquisition and processing: A) Measurement from the highest point of the furcation area to 2 mm apically; B) Measurement of the risk zone before cervical preparation; C) Measurement of the risk zone after cervical preparation.

Results

The values for dentin thickness in the risk zone before and after cervical flaring with Gates-Gliden drills were 1.1373 (± 0.2851) mm and 0.7573 (± 0.2663), respectively. There was statistically significant difference between groups. The values showed a small dentin thickness after cervical flaring with Gates-Gliden drills (Figure 3).

Discussion

CBCT allows evaluating root canal preparation without cutting off the specimens and losing the root canal structure




Fig. 3. Mean values and standard deviations for the distances between the root canal center and the furcation area for Group 1 (before Gates-Gliden preparation) and for Group 2 (after Gattes-Glidden preparation) samples.

material during sectioning¹⁸⁻²³. The methodology proposed in this study permits standardizing specimen positioning for acquisition of CBCT images before and after cervical flaring. We used squared gypsum pattern in order to align the samples through the horizontal and vertical bars of the software IcatVision®. The images before and after preparation were superimposed. In addition, we used two computers to evaluate the pre- and post-instrumentation images, which were standardized and measured in the software. Thus, there were no other image manipulation programs for generating distortion attempt to superimpose images of the same size before and after preparation. In addition, the “zoom” tool was employed in order to best measure the specimens.

Dentin thickness in the risk zone after the action of instruments for cervical flaring can be evaluated by different methodologies. Methods that use cross-sections of the sample do not provide an adequate stabilization for the tooth structure¹⁶. This study suggested a standardizing methodology for acquisition of CBCT images. Assessment of the risk zone could be carried out by micro-computed tomography (μ CT), but it is time-consuming, expensive and not always available in all research centers.

Several methods have been proposed for standardization of imaging acquisition. Bramante (1987)²⁴ suggested a method in which images obtained from root slices were superimposed to detect alterations that were produced after root canal preparation. Zaia et al. (2000)²⁵, Wu (2005)²⁶ and Sauáia et al. (2010)²³ modified the previously described methodology and suggested to include the teeth in acrylic resin. Coutinho Filho et al. (2008)¹⁴ used a plastic tube and a metallic strip to guide the slices repositioning. Our study also adopted this protocol in order to standardize the position of each section to evaluate the same sample before and after root canal preparation. However, the squared-shaped samples inside acrylic resin blocks were placed in a box with plaster to guarantee their perfect repositioning. CBCT was also adopted to avoid sample sectioning, ensuring that a significant loss of dental structure did not occur.

It is feasible to conclude that the use of the present methodology allows comparing specimens before and after

root canal instrumentation, since image superimposition was obtained after acquisition of CBCT images.

References

- Weine FS, Pasiewicz RA, Rice RT. Canal configuration of the mandibular second molar using a clinically oriented in vitro method. *J Endod.* 1988; 14: 207-13.
- Leeb J. Canal orifice enlargement as related to biomechanical preparation. *J Endod.* 1983; 9: 463-70.
- Schilder H. Cleaning and shaping the root canal. *Dent Clin North Am.* 1974; 18: 269-96.
- Abou-Rass M, Frank AL, Glick DH. The anticurvature filing method to prepare the curved root canal. *J Am Dent Assoc.* 1980; 101: 792-94.
- Kessler JR, Peters DD, Lorton L. Comparison of the relative risk of molar root perforations using various endodontic instrumentation techniques. *J Endod.* 1983; 9: 439-47.
- Montgomery S. Root canal wall thickness of mandibular molars after biomechanical preparation. *J Endod.* 1985; 11: 6257-63.
- Abou-rass M, Jastrab RJ. The use of rotary instruments as auxiliary AIDS to root canal preparation of molars. *J Endod.* 1982; 8: 78-82.
- McCann JT, Keller DL, Labounty GL. A modification of the muffle model system to study root canal morphology. *J Endod.* 1990; 16: 114-5.
- Isom TL, Marshall JG, Baumgartner JC. Evaluation of root thickness in curved canals after flaring. *J Endod.* 1995; 21: 368-71.
- Coutinho-Filho T, de Deus G, Pinto TG, Gurgel-Filho ED, Maniglia-Ferreira C. A computer evaluation of the dentin remaining after cervical preparation in curved canals: Gates-glidden drills vs. orifice shaper. *Braz J Oral Sci.* 2002; 1: 116-20.
- Zuckerman O, Katz A, Pilo R, Tamse A, Fuss Z. Residual dentin thickness in mesial roots of mandibular molars prepared with Lightspeed rotary instruments and Gates-Glidden reamers. *Oral Surg, Oral Med, Oral Pathol, Oral Radiol and Endod.* 2003; 96: 351-55.
- Wu MK, Van der Sluis LWM, Wesselink PR. The risk of furcal perforation in mandibular molars using Gates-Glidden drills with anticurvature pressure. *Oral Surg, Oral Med, Oral Pathol, Oral Radiol and Endod.* 2005; 99: 378-82.
- Constante IGT, Davidowicz H, Barletta FB, de Moura AAM. Study of the areas and thicknesses of mesiobuccal root canals prepared by three endodontic techniques. *Braz Oral Res.* 2007; 21: 118-26.
- Coutinho-filho T, de Deus G, Gurgel-Filho ED, Rocha-Lima AC, Dias KRC, Barbosa CA. Evaluation of the risk of a stripping perforation with gates-glidden drills: serial versus crown-down sequences. *Braz Oral Res.* 2008; 22: 18-24.
- Kuttler S, Garala M, Perez R, Dorn SO. The endodontic cube: a system designed for evaluation of root canal anatomy and canal preparation. *J Endod.* 2001; 27: 533-6.
- Garala M, Kuttler S, Hardigan P, Steiner-Carmi R, Dorn S. A comparison of the minimum canal wall thickness remaining following preparation using two nickel-titanium rotary systems. *Int Endod J.* 2003; 36: 636-42.
- Mahrar AH, Aboel-Fotouh MM. Comparison of effects of protaper, heroshaper, and gates glidden burs on cervical dentin thickness and root canal volume by using multislice computed tomography. *J Endod.* 2008; 34: 1219-22.
- Sanfelice CM, da Costa FB, Só MVR, Vier-Pelisser F, Bier CAS, Grecca FS. Effects of Four Instruments on Coronal Pre-enlargement by Using Cone Beam Computed Tomography. *J Endod.* 2010; 36: 858-61.
- Cotton TP, Geisler TM, Holden DT, Schwartz AS, Schindler WG. Endodontic application of cone beam volumetric tomography. *J Endod.* 2007; 9: 1121-32.
- Patel S, Dawood A, Ford TP, Whaites E. The potential applications of cone beam computed tomography in the management of endodontic problems. *Int Endod J.* 2007; 40: 818-30.
- Patel S, Dawood A, Whaites E, Ford TP. New dimensions in endodontic imaging: part 1. Conventional and alternative radiographic systems. *Int Endod J.* 2009; 42: 447-62.
- Michetti J, Maret D, Mallet JP, Diemer F. Validation of cone beam computed tomography as a tool to explore root canal anatomy. *J Endod.* 2010; 36: 1187-90.
- Sauáia TS, Gomes BPFA, Pinheiro ET, Zaia A A, Ferraz CCR, Souza-Filho FJ, Valdrighi L. Thickness of dentine in mesial roots of mandibular molars with different lengths. *Int Endod J.* 2010; 43: 555-9.
- Bramante CM, Berbert A, Borges RP. A Methodology for Evaluation of Root Canal Instrumentation. *J Endod.* 1987; 13: 223-5.
- Zaia AA, Ferraz CCR, Yoshinari GH, Souza Filho FJ. A Simple Method for the Analysis of Root Canal Preparation. *J Endod.* 2000; 26: 172-4.
- Wu MK, Van der Sluis WM, Wesselink PR. The risk of furcal perforation in mandibular molars using Gates-Glidden drills with anticurvature pressure. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2005; 99: 378-82.